




# Certificate of Calibration

Pile Dynamics, Inc.  
 Transducer Model: PDI Strain Transducer  
 Serial Number: C258  
 PDI Gage Factor: 91.0  
 Mean Linear Correlation Coefficient: 0.999906

Table 1: Representative Calibration Data

Applied strain ( $\mu E$ )	Gage Output (mV/Vexc)	Applied strain ( $\mu E$ )	Gage Output (mV/Vexc)
-2080.1	-6.520	629.9	2.103
-1548.6	-4.896	984.3	3.218
-1076.1	-3.408	1329.8	4.330
-656.2	-2.014	1699.5	5.440
-218.7	-0.603	2040.7	6.469
229.7	0.824		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: Paul Hartman

PDI Gage: C258

Calibration Date: 9/7/2023 7:06:05 AM

