




# Certificate of Calibration

Pile Dynamics, Inc.  
 Transducer Model: PDI Strain Transducer  
 Serial Number: F372  
 PDI Gage Factor: 89.0  
 Mean Linear Correlation Coefficient: 0.999826

Table 1: Representative Calibration Data

Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)	Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)
-1863.5	-5.938	531.5	1.883
-1371.4	-4.386	866.1	2.988
-958.0	-3.035	1227.0	4.134
-577.4	-1.787	1581.4	5.221
-177.2	-0.492	1988.2	6.438
203.4	0.787		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: \_\_\_\_\_

*Paul Hartman*

PDI Gage: F372

Calibration Date: 6/7/2023 12:55:11 PM

