




# Certificate of Calibration

Pile Dynamics, Inc.  
 Transducer Model: PDI Strain Transducer  
 Serial Number: P800  
 PDI Gage Factor: 137.8  
 Mean Linear Correlation Coefficient: 0.999821

Table 1: Representative Calibration Data

Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)	Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)
-1797.9	-3.801	610.2	1.306
-1332.0	-2.839	951.4	2.027
-971.1	-2.024	1312.3	2.777
-584.0	-1.192	1706.0	3.546
-170.6	-0.323	1975.1	4.086
249.3	0.546		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: \_\_\_\_\_

*Paul Hartman*

PDI Gage: P800

Calibration Date: 6/2/2023 5:07:07 AM

