




Certificate of Calibration

Pile Dynamics, Inc.
 Transducer Model: PDI Strain Transducer
 Serial Number: P821
 PDI Gage Factor: 138.1
 Mean Linear Correlation Coefficient: 0.999972

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-2073.5	-4.308	643.0	1.360
-1542.0	-3.232	997.4	2.112
-1069.6	-2.251	1351.7	2.852
-656.2	-1.355	1717.0	3.596
-223.1	-0.444	1988.2	4.154
223.1	0.490		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: Paul Hartman

PDI Gage: P821

Calibration Date: 9/20/2023 6:30:51 AM

