




Certificate of Calibration

TechSense
 Transducer Model: PDI Strain Transducer
 Serial Number: V414
 PDI Gage Factor: 140.2
 Mean Linear Correlation Coefficient: 0.999749

Table 1: Representative Calibration Data

| Applied strain (μE) | Gage Output (mV/Vexc) | Applied strain (μE) | Gage Output (mV/Vexc) |
|----------------------------------|-----------------------|----------------------------------|-----------------------|
| -1852.3 | -4.057 | 64.6 | 0.062 |
| -1316.6 | -2.969 | 555.4 | 1.149 |
| -886.4 | -2.047 | 970.5 | 2.029 |
| -520.6 | -1.243 | 1386.5 | 2.883 |
| -169.7 | -0.465 | 1723.8 | 3.552 |

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

| | |
|-------------------------------------|-------------------|
| PDI STCS Serial Number: | 1003HA |
| Firmware version number: | 0.9.0.0 |
| Transducer Gage Length: | 3 inches (76.2mm) |
| Excitation Voltage for Calibration: | 5.0 VDC |

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: RASENDU PATEL

Signature: _____

PDI Gage: V414

Calibration Date: 13-02-2023 15:17:54



TechSense

49, Atmajyoti Nagar Society, Near Atmajyoti Ashram, Ellora Park, Vadodara - 390023
 Website : www.techsenseindia.com E-mail : info@techsenseindia.com
 Mobile.: +91- 8905510013 / +91 9327226069