




Certificate of Calibration

Pile Dynamics, Inc.
Transducer Model: PDI Strain Transducer
Serial Number: W335
PDI Gage Factor: 93.4
Mean Linear Correlation Coefficient: 0.999832

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-1850.0	-5.541	441.7	1.567
-1341.7	-3.972	750.0	2.568
-825.0	-2.409	1091.7	3.605
-325.0	-0.871	1425.0	4.614
108.3	0.478	1958.3	6.142

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: _____

Paul Hartman

PDI Gage: W335

Calibration Date: 4/2/2024 5:53:54 AM

