




# Certificate of Calibration

Pile Dynamics, Inc.  
 Transducer Model: PDI Strain Transducer  
 Serial Number: U131  
 PDI Gage Factor: 141.4  
 Mean Linear Correlation Coefficient: 0.999978

Table 1: Representative Calibration Data

Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)	Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)
-2316.3	-4.668	433.1	0.964
-1874.5	-3.743	892.4	1.910
-1430.4	-2.828	1299.2	2.757
-977.7	-1.899	1741.0	3.644
-524.9	-0.964	2152.2	4.462
-59.1	-0.036		

Calibration Curve




PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: 

PDI Gage: U131

Calibration Date: 4/25/2024 9:47:44 AM

