




Certificate of Calibration

Pile Dynamics, Inc.
 Transducer Model: PDI Strain Transducer
 Serial Number: T409
 PDI Gage Factor: 142.0
 Mean Linear Correlation Coefficient: 0.999921

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-2248.5	-4.537	459.3	1.010
-1725.7	-3.488	794.0	1.688
-1233.6	-2.485	1128.6	2.370
-774.3	-1.543	1496.1	3.091
-328.1	-0.616	1857.0	3.784
105.0	0.271	1981.6	4.034

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: _____

Paul Hartman

PDI Gage: T409

Calibration Date: 6/17/2024 7:04:29 AM

