



Certificate of Calibration

Pile Dynamics, Inc.

Transducer Model: PDI Strain Transducer

Serial Number: Z914


PDI Gage Factor: 97.1

Mean Linear Correlation Coefficient: 0.999976

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-2965.9	-8.561	474.6	1.721
-2276.9	-6.478	977.7	3.229
-1607.6	-4.470	1410.8	4.497
-964.6	-2.587	1837.3	5.720
-393.7	-0.898	1981.6	6.151
0.0	0.306		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: VANNA OUK

Signature: _____

PDI Gage: Z914

Calibration Date: MAY 14 2024

