

# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.  
 Calibration performed on

**JUL 18 2024**

Serial No: 80162      Temperature: 24.2 °C

Model: PE      Humidity: 40%

Calibrated on: Channel 3 on 8G 5161 LE

**PDA CALIBRATION FACTOR**

**897.0 g's/volt**

R<sup>2</sup>: 0.999944 [Chip programmed]

Operator: William Johnson

Ref Acc 1: 78268!      Cal on: 11Jan2024  
 986 g's/volt


Ref Acc 2: 78270!      Cal on: 11Jan2024  
 971 g's/volt

*William Johnson*

Signed



Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N 80162 Velocity
m/s	m/s
1.312	1.306
1.666	1.667
1.948	1.953
2.273	2.266
2.574	2.583
2.867	2.859
3.132	3.124
3.439	3.438
3.732	3.733
4.045	4.055

Maximum Acceleration: 873 g's

# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.  
Calibration performed on

JUL 18 2024

Serial No: 80014      Temperature: 23.5 °C

Model: PE      Humidity: 41%

Calibrated on: Channel 3 on 8G 5161 LE

PDA CALIBRATION FACTOR

914.9 g's/volt

R<sup>2</sup>: 0.999944 [Chip programmed]

Operator: William Johnson

Ref Acc 1: 78268!      Cal on: 11Jan2024  
986 g's/volt


Ref Acc 2: 78270!      Cal on: 11Jan2024  
971 g's/volt

*William Johnson*

Signed



Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N 80014 Velocity
m/s	m/s
1.357	1.351
1.651	1.646
1.972	1.971
2.282	2.273
2.577	2.573
2.872	2.886
3.143	3.148
3.447	3.449
3.745	3.747
4.013	4.007

Maximum Acceleration: 864 g's

# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.  
 Calibration performed on **JUL 19 2024**

Serial No: K13859      Temperature: 21.9 °C  
 Model: PR      Humidity: 27%  
 Calibrated on: Channel 3 on 8G 5161 LE

### PDA CALIBRATION FACTOR

**365.4 mv/5000g**  
 (73.1  $\mu\text{v/g}$ )  
 R<sup>2</sup>: 0.999953 [Chip programmed]


Ref Acc 1: 72517!      Cal on: 24Mar2022  
 1049 g's/volt  
 Ref Acc 2: 72505!      Cal on: 24Mar2022  
 1035 g's/volt

Operator: William Johnson

Signed



Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N K13859 Velocity
m/s	m/s
1.068	1.063
1.517	1.522
1.839	1.828
2.262	2.256
2.618	2.613
2.907	2.919
3.177	3.182
3.478	3.478
3.777	3.782
4.127	4.121

Maximum Acceleration: 884 g's

# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.  
 Calibration performed on

**JUL 19 2024**

**Serial No:** K14032      **Temperature:** 24.9 °C  
**Model:** PR      **Humidity:** 40%  
**Calibrated on:** Channel 4 on 8G 5161 LE

### PDA CALIBRATION FACTOR

**349.8 mv/5000g**  
 (70.0  $\mu\text{v/g}$ )  
 R<sup>2</sup>: 0.999964 [Chip programmed]


**Ref Acc 1:** 78270!      **Cal on:** 11Jan2024  
 971 g's/volt  
**Ref Acc 2:** 78268!      **Cal on:** 11Jan2024  
 986 g's/volt

Operator: William Johnson

Signed



Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N K14032 Velocity
m/s	m/s
1.054	1.054
1.526	1.526
1.912	1.908
2.225	2.219
2.481	2.479
2.750	2.759
3.059	3.053
3.385	3.378
3.690	3.699
3.979	3.982

Maximum Acceleration: 859 g's

# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.  
 Calibration performed on **JUL 19 2024**

**Serial No:** K14033      **Temperature:** 24.9 °C  
**Model:** PR      **Humidity:** 40%  
**Calibrated on:** Channel 3 on 8G 5161 LE

### PDA CALIBRATION FACTOR

**355.6 mv/5000g**  
 (71.1  $\mu\text{v/g}$ )  
 R<sup>2</sup>: 0.999950 [Chip programmed]


**Ref Acc 1:** 78270!      **Cal on:** 11Jan2024  
 971 g's/volt  
**Ref Acc 2:** 78268!      **Cal on:** 11Jan2024  
 986 g's/volt

Operator: William Johnson

Signed



Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N K14033 Velocity
m/s	m/s
1.054	1.050
1.526	1.528
1.912	1.923
2.225	2.235
2.481	2.482
2.750	2.742
3.059	3.060
3.385	3.387
3.690	3.682
3.979	3.978

Maximum Acceleration: 859 g's

# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.  
Calibration performed on

JUL 19 2024

Serial No: K14034      Temperature: 24.9 °C

Model: PR      Humidity: 40%

Calibrated on: Channel 4 on 8G 5161 LE

### PDA CALIBRATION FACTOR

**322.3 mv/5000g**  
(64.5  $\mu\text{v/g}$ )  
R<sup>2</sup>: 0.999929 [Chip programmed]

Operator: William Johnson


Ref Acc 1: 78270!      Cal on: 11Jan2024  
971 g's/volt

Ref Acc 2: 78268!      Cal on: 11Jan2024  
986 g's/volt

Signed



Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N K14034 Velocity
m/s	m/s
1.085	1.090
1.511	1.499
1.884	1.890
2.217	2.224
2.483	2.498
2.784	2.779
3.083	3.080
3.389	3.380
3.684	3.684
3.993	3.994

Maximum Acceleration: 862 g's

# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.  
Calibration performed on

JUL 19 2024

Serial No: K14035      Temperature: 24.9 °C  
Model: PR      Humidity: 40%  
Calibrated on: Channel 3 on 8G 5161 LE

### PDA CALIBRATION FACTOR

**334.8 mv/5000g**  
(67.0  $\mu\text{v/g}$ )  
R<sup>2</sup>: 0.999929 [Chip programmed]

Operator: William Johnson


Ref Acc 1: 78270!      Cal on: 11Jan2024  
971 g's/volt  
Ref Acc 2: 78268!      Cal on: 11Jan2024  
986 g's/volt

*William Johnson*

Signed



Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N K14035 Velocity
m/s	m/s
1.085	1.084
1.511	1.523
1.884	1.892
2.217	2.219
2.483	2.477
2.784	2.778
3.083	3.082
3.389	3.401
3.684	3.675
3.993	3.991

Maximum Acceleration: 862 g's

# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.  
 Calibration performed on

JUL 19 2024

Serial No: K14065      Temperature: 25.7 °C  
 Model: PR      Humidity: 42%  
 Calibrated on: Channel 3 on 8G 5161 LE

### PDA CALIBRATION FACTOR


**346.4 mv/5000g**  
 (69.3  $\mu$ v/g)  
 R<sup>2</sup>: 0.999964 [Chip programmed]

Ref Acc 1: 78268!      Cal on: 11Jan2024  
 986 g's/volt  
 Ref Acc 2: 78270!      Cal on: 11Jan2024  
 971 g's/volt

Operator: William Johnson

*William Johnson*  
 Signed

Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N K14065 Velocity
m/s	m/s
1.129	1.137
1.641	1.648
2.032	2.035
2.399	2.400
2.788	2.796
3.075	3.076
3.420	3.424
3.737	3.726
4.028	4.031
4.350	4.343

Maximum Acceleration: 945 g's






# Certificate of Calibration

Pile Dynamics, Inc.  
 Transducer Model: PDI Strain Transducer  
 Serial Number: 38A  
 PDI Gage Factor: 93.3  
 Mean Linear Correlation Coefficient: 0.999965

Table 1: Representative Calibration Data

Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)	Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)
-2073.5	-6.292	374.0	1.321
-1568.2	-4.734	794.0	2.598
-1091.4	-3.241	1168.0	3.780
-689.0	-2.008	1502.6	4.801
-350.0	-0.970	1824.1	5.773
-26.2	0.063	1981.6	6.210

Calibration Curve




PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: 

PDI Gage: 38A

Calibration Date:                     JUL 19 2024                    






# Certificate of Calibration

Pile Dynamics, Inc.  
 Transducer Model: PDI Strain Transducer  
 Serial Number: Z956  
 PDI Gage Factor: 93.6  
 Mean Linear Correlation Coefficient: 0.999990

Table 1: Representative Calibration Data

Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)	Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)
-2053.8	-6.239	406.8	1.358
-1542.0	-4.671	826.8	2.664
-1069.6	-3.173	1213.9	3.861
-656.2	-1.923	1548.6	4.890
-334.6	-0.937	1863.5	5.850
-4.4	0.080	1968.5	6.155

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: VANNA OUK

Signature: 

PDI Gage: Z956

Calibration Date: JUL 18 2024






# Certificate of Calibration

Pile Dynamics, Inc.  
 Transducer Model: PDI Strain Transducer  
 Serial Number: Z957  
 PDI Gage Factor: 92.3  
 Mean Linear Correlation Coefficient: 0.999972

Table 1: Representative Calibration Data

Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)	Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)
-2007.9	-6.236	387.1	1.253
-1509.2	-4.690	780.8	2.517
-1032.4	-3.217	1122.0	3.576
-675.9	-2.090	1443.6	4.560
-367.5	-1.114	1752.0	5.507
-26.2	-0.059	1972.9	6.170

Calibration Curve




PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: VANNA OUK

Signature: 

PDI Gage: Z957

Calibration Date: JUL 18 2024



# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.

Calibration performed on **JUL 19 2024**

Serial No: 79618      Temperature: 23.6 °C

Model: PE      Humidity: 54%

Calibrated on: Channel 3 on 8G 5161 LE

**PDA CALIBRATION FACTOR**

**933.8 g's/volt**

R<sup>2</sup>: 0.999959 [No chip]

Operator: William Johnson


Ref Acc 1: 78270!      Cal on: 11Jan2024  
971 g's/volt

Ref Acc 2: 78268!      Cal on: 11Jan2024  
986 g's/volt

Signed



Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N 79618 Velocity
m/s	m/s
0.992	0.995
1.447	1.441
1.795	1.802
2.098	2.105
2.449	2.452
2.777	2.785
3.102	3.103
3.393	3.387
3.664	3.664
3.956	3.946

Maximum Acceleration: 845 g's




# Certificate of Calibration

Pile Dynamics, Inc.  
 Transducer Model: PDI Strain Transducer  
 Serial Number: Y916  
 PDI Gage Factor: 92.6  
 Mean Linear Correlation Coefficient: 0.999776

Table 1: Representative Calibration Data

Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)	Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)
-1902.9	-6.013	465.9	1.521
-1397.6	-4.401	833.3	2.651
-925.2	-2.892	1128.6	3.575
-479.0	-1.506	1437.0	4.464
-183.7	-0.536	1730.1	5.338
111.5	0.390	2014.4	6.169

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Kay Tol

Signature: \_\_\_\_\_

*Kay Tol*

PDI Gage: Y916

Calibration Date: \_\_\_\_\_

JUL 19 2024



# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.  
Calibration performed on

**JUL 18 2024**

Serial No: K13941      Temperature: 24.7 °C  
Model: PR      Humidity: 47%  
Calibrated on: Channel 3 on 8G 5161 LE

### PDA CALIBRATION FACTOR

**389.2 mv/5000g**  
(77.8  $\mu\text{v/g}$ )  
R<sup>2</sup>: 0.999914 [Chip programmed]


Ref Acc 1: 78270!      Cal on: 11Jan2024  
971 g's/volt  
Ref Acc 2: 78268!      Cal on: 11Jan2024  
986 g's/volt

Operator: William Johnson

Signed



Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N K13941 Velocity
m/s	m/s
1.004	0.996
1.532	1.541
1.942	1.921
2.305	2.292
2.736	2.730
3.085	3.083
3.436	3.447
3.723	3.728
4.013	4.018
4.317	4.320

Maximum Acceleration: 935 g's

# Accelerometer Calibration Certificate

## Pile Dynamics, Inc.



Calibrated by Pile Dynamics, Inc.  
Calibration performed on

JUL 18 2024

Serial No: K13939      Temperature: 24.7 °C  
Model: PR      Humidity: 48%  
Calibrated on: Channel 3 on 8G 5161 LE

### PDA CALIBRATION FACTOR

**389.4 mv/5000g**  
(77.9  $\mu$ v/g)  
R<sup>2</sup>: 0.999917 [Chip programmed]


Ref Acc 1: 78270!      Cal on: 11Jan2024  
971 g's/volt  
Ref Acc 2: 78268!      Cal on: 11Jan2024  
986 g's/volt

Operator: William Johnson

Signed



Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).



Reference Velocity	S/N K13939 Velocity
m/s	m/s
1.036	1.037
1.542	1.532
1.964	1.947
2.382	2.366
2.730	2.730
3.074	3.075
3.429	3.440
3.722	3.733
3.999	3.998
4.317	4.320

Maximum Acceleration: 936 g's




# Certificate of Calibration

Pile Dynamics, Inc.  
 Transducer Model: PDI Strain Transducer  
 Serial Number: 33A  
 PDI Gage Factor: 94.1  
 Mean Linear Correlation Coefficient: 0.999946

Table 1: Representative Calibration Data

Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)	Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)
-2064.7	-6.263	367.5	1.234
-1561.7	-4.722	774.3	2.497
-1076.1	-3.245	1154.9	3.661
-643.0	-1.921	1489.5	4.661
-336.8	-0.954	1797.9	5.584
-13.1	0.042	1968.5	6.054

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature:

PDI Gage: 33A

JUL 18 2024

Calibration Date: \_\_\_\_\_






# Certificate of Calibration

Pile Dynamics, Inc.  
 Transducer Model: PDI Strain Transducer  
 Serial Number: 34A  
 PDI Gage Factor: 94.6  
 Mean Linear Correlation Coefficient: 0.999981

Table 1: Representative Calibration Data

Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)	Applied strain ( $\mu\text{E}$ )	Gage Output (mV/Vexc)
-2086.6	-6.260	360.9	1.236
-1594.5	-4.726	780.8	2.518
-1095.8	-3.251	1181.1	3.723
-669.3	-1.933	1528.9	4.791
-354.3	-0.965	1870.1	5.833
-13.1	0.042	1981.6	6.144

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: \_\_\_\_\_

*Paul Hartman*

PDI Gage: 34A

**JUL 18 2024**

Calibration Date: \_\_\_\_\_

