


Cycle 1
Sample

Force (kips)	Bridge (mV/V)
1 0	0
2 0.9506901	-0.05398548
3 2.12040994	-0.120979274
4 3.09966014	-0.175770451
5 4.1519201	-0.236148843
6 5.24200078	-0.297512984
7 6.00933638	-0.34101621
8 7.17203684	-0.406782914
9 7.94707295	-0.450788548
10 9.16163082	-0.519659381
11 9.9399811	-0.564044225

Serial Number 30389
Force Calibration (lb/V) -17.6394922
Offset -0.00640346
Correlation -0.99999889




Cycle 2
Sample

Force (kips)	Bridge (mV/V)
1 0	0
2 1.12478344	-0.062207462
3 2.0001238	-0.110691121
4 2.99126619	-0.166951349
5 4.0688701	-0.226481467


6	5.14666876	-0.288082542
7	6.31385326	-0.353474811
8	7.0877197	-0.396854865
9	8.21230674	-0.459963094
10	9.00020973	-0.504437924
11	10.0456458	-0.570290125

Serial Number 30389
 Force Calibration (kips/mV/V) -17.6737489
 Offset 0.0505866
 Correlation -0.99993303



Sample	Cycle 3	
	Force (kips)	Bridge (mV/V)
1	0	0
2	1.13414117	-0.063757085
3	1.99378777	-0.11206067
4	3.20464243	-0.18013025
5	3.89916325	-0.219268867
6	4.99041369	-0.280291793
7	6.13079214	-0.345684058
8	6.91645325	-0.389760732
9	8.12798939	-0.457152947
10	8.90400024	-0.503684474
11	10.1929327	-0.575441394

Serial Number 30389
Force Calibration (kips/mV/V) -17.6946711
Offset 0.01589696
Correlation -0.99999074



Serial Number 30389
Force Calibration (kips/mV/V) -17.6693041
Squid Gain 50
Squid Excitation 5
Force Calibration (kips/V) -3533.86081

Calibrated by: Sean Bonney
Calibrated Date: 9/3/2024

Pile Dynamics Inc
30725 Aurora Rd
Solon, OH 44139

Traceable to N.I.S.T.

