


Cycle 1
Sample

Force (kips)	Bridge (mV/V)
1	0
2	0.94562125
3	1.83704522
4	2.86279178
5	3.90803343
6	5.0235555
7	5.80853436
8	6.95251932
9	7.75172962
10	8.95459029
11	9.67289502

Serial Number 30390
Force Calibration (lb/V) -17.6684614
Offset 0.0025966
Correlation -0.99999743




Cycle 2
Sample

Force (kips)	Bridge (mV/V)
1	0
2	0.88177407
3	2.11027232
4	3.16526183
5	3.87606125


6	4.98534485	-0.282873764
7	5.95465192	-0.337366354
8	7.00476695	-0.397166557
9	8.18657273	-0.463468594
10	9.00186663	-0.510213205
11	10.1235293	-0.573501408

Serial Number	30390
Force Calibration (kips/mV/V)	-17.6994978
Offset	-0.02395966
Correlation	-0.99999778



Sample	Cycle 3	
	Force (kips)	Bridge (mV/V)
1	0	0
2	1.1425242	-0.06519282
3	2.01464782	-0.114557795
4	3.25630511	-0.185304666
5	4.33488373	-0.245768192
6	5.09578598	-0.288740494
7	6.23304518	-0.352554499
8	7.02348264	-0.397441378
9	8.22244431	-0.464937611
10	9.15383278	-0.523121951
11	10.252686	-0.586703972

Serial Number 30390
Force Calibration (kips/mV/V) -17.5253106
Offset 0.02223692
Correlation -0.99993934



Serial Number 30390
Force Calibration (kips/mV/V) -17.6310899

Squid Gain 50
Squid Excitation 5
Force Calibration (kips/V) -3526.21799

Calibrated by: Sam Boney
Calibrated Date: 9/4/2024

Pile Dynamics Inc
30725 Aurora Rd
Solon, OH 44139

Traceable to N.I.S.T.

