


Cycle 1
Sample

Force (kips)	Bridge (mV/V)
1 0	0
2 1.13063205	-0.06413145
3 2.02917184	-0.11505564
4 2.99994161	-0.169586168
5 4.05229908	-0.228410187
6 5.15222493	-0.291589316
7 6.28168599	-0.354882243
8 7.04726693	-0.398030088
9 8.24086738	-0.466280046
10 9.01190692	-0.510584276
11 10.1572564	-0.579251405

Serial Number 30392
Force Calibration (lb/V) -17.5922671
Offset 0.01985324
Correlation -0.99997056




Cycle 2
Sample

Force (kips)	Bridge (mV/V)
1 0	0
2 1.15139455	-0.06489442
3 2.00957899	-0.113946751
4 2.98502762	-0.168429896
5 4.01798724	-0.227860502


6	5.13536138	-0.291655699
7	5.89012251	-0.334722944
8	7.05506495	-0.402603201
9	8.2428168	-0.471819918
10	9.04524379	-0.519308722
11	10.2658452	-0.595600827

Serial Number 30392
 Force Calibration (kips/mV/V) -17.257278
 Offset 0.07331802
 Correlation -0.99991373



Sample	Cycle 3	
	Force (kips)	Bridge (mV/V)
1	0	0
2	0.84872945	-0.048896013
3	1.97507217	-0.11296605
4	2.90051527	-0.165748057
5	3.91271235	-0.223297462
6	5.00435267	-0.285457899
7	5.75043838	-0.328274098
8	6.90943474	-0.393619221
9	8.11239293	-0.461665625
10	8.89473975	-0.506657151
11	9.96893138	-0.576191751

Serial Number 30392
Force Calibration (kips/mV/V) -17.4403379
Offset 0.01639038
Correlation -0.99991298



Serial Number 30392
Force Calibration (kips/mV/V) -17.429961
Squid Gain 50
Squid Excitation 5
Force Calibration (kips/V) -3485.9922

Calibrated by: Sean Bonney
Calibrated Date: 9/3/2024

Pile Dynamics Inc
30725 Aurora Rd
Solon, OH 44139

Traceable to N.I.S.T.

