


Cycle 1
Sample

	Force (kips)	Bridge (mV/V)
1	0	0
2	0.80525486	-0.046365504
3	1.98871891	-0.113260031
4	3.0196317	-0.171051164
5	4.06721277	-0.231226039
6	4.82830998	-0.273720024
7	5.95250732	-0.337709862
8	7.11559768	-0.403917612
9	7.90866695	-0.449013431
10	9.10840838	-0.518557568
11	9.94582945	-0.568458814

Serial Number 30392
Force Calibration (lb/V) -17.5453111
Offset 0.0116671
Correlation -0.99998097




Cycle 2
Sample

	Force (kips)	Bridge (mV/V)
1	0	0
2	0.90019713	-0.050677957
3	1.76003869	-0.099668902
4	2.75498266	-0.155123785
5	3.81426097	-0.21547404


6	4.91720858	-0.276947495
7	6.06285076	-0.341553469
8	6.85270336	-0.385137554
9	8.03061008	-0.452634443
10	8.85418948	-0.498507556
11	9.98491751	-0.566464304

Serial Number	30392
Force Calibration (kips/mV/V)	-17.6985982
Offset	0.00894695
Correlation	-0.99997537



Sample	Cycle 3	
	Force (kips)	Bridge (mV/V)
1	0	0
2	0.87826491	-0.049787058
3	2.04135634	-0.116018223
4	3.03756747	-0.171804868
5	4.08436873	-0.229942084
6	5.19550435	-0.29397461
7	5.97726648	-0.337549192
8	7.12758741	-0.402657607
9	7.93284125	-0.447601852
10	9.12663661	-0.515463771
11	9.94924125	-0.56145545

Serial Number 30392
Force Calibration (kips/mV/V) -17.7285464
Offset -0.0073317
Correlation -0.9999973



Serial Number 30392
Force Calibration (kips/mV/V) -17.6574852

Squid Gain 50
Squid Excitation 5
Force Calibration (kips/V) -3531.49704

Calibrated by: Sean Bonner
Calibrated Date: 7/31/2024

Pile Dynamics Inc
30725 Aurora Rd
Solon, OH 44139

Traceable to N.I.S.T.

