



Hydrophone Calibration Data

Model Number: **H507A164**

Serial Number: **10056**

Crystal Sensitivity: **-193.0 dB**
(re 1 V/ μ Pa)

Midband Gain: **10.2 dB**

Bias Voltage: **9.97 Vdc**

Calibrated by: **T.PHOUBANDITH** Date: **9/24/2024**

This calibration is traceable to the National Institute of Standards and Technology, Gaithersburg, MD 20899.

Internal Cal Attenuation = -14dB; 40 Ohm & 10 Ohm, 1%, 0.1 W Resistors

CAUTION: 1Vac rms maximum calibration input

Wilcoxon Sensing Technologies is an ISO 9001 Registered Company.