
Certificate of Calibration

Pile Dynamics, Inc.
Transducer Model: PDI Strain Transducer
Serial Number: 139A
PDI Gage Factor: 93.6
Mean Linear Correlation Coefficient: 0.999918

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-1889.8	-5.809	557.7	1.807
-1386.7	-4.247	977.7	3.105
-905.5	-2.759	1345.1	4.241
-570.9	-1.681	1692.9	5.261
-229.7	-0.643	1981.6	6.095
124.7	0.465		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman Signature: Paul Hartman

PDI Gage: 139A
Calibration Date: SEP 2 0 2024

Pile Dynamics, Inc.