Certificate of Calibration
Pile Dynamics, Inc.

Calibrated by Pile Dynamics, Inc.

Serial No:	80157	**Temperature:**	24.6 °C
Model:	PE	**Humidity:**	49%
Calibrated on:	Channel 4 on 8G 5161 LE		

Ref Acc 1:	78270! 971 g's/volt	**Cal on:**	11Jan2024
Ref Acc 2:	78268! 986 g's/volt	**Cal on:**	11Jan2024

Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).

PDA CALIBRATION FACTOR

970.6 g's/volt

R^2: 0.999979 [Chip programmed]

Operator: William Johnson

Signed

Calibration valid 2 years

From: SEP 2 5 2024



Reference Velocity	S/N 80157 Velocity
m/s	m/s
0.969	0.965
1.357	1.351
1.695	1.697
2.069	2.072
2.429	2.422
2.786	2.791
3.161	3.159
3.557	3.554
3.884	3.881
4.316	4.324

Maximum Acceleration: 928 g's

PILE DYNAMICS, INC
ORIGINAL COPY

Version: 2024.35.11 0.18