



Certificate of Calibration

Pile Dynamics, Inc.

Transducer Model: PDI Strain Transducer

Serial Number: K014


PDI Gage Factor: 90.2

Mean Linear Correlation Coefficient: 0.999712

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-2106.3	-6.830	183.7	0.643
-1581.4	-5.184	590.6	1.943
-1194.2	-3.883	977.7	3.150
-813.6	-2.632	1345.1	4.279
-498.7	-1.564	1699.5	5.353
-170.6	-0.525	2007.9	6.262

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature:

PDI Gage: K014

Calibration Date: 9/6/2024 5:51:37 AM

