

Certificate of Calibration

Pile Dynamics, Inc.

Calibrated by Pile Dynamics, Inc.



Serial No: LW191303 Temperature: 21.8 °C

Model: PE Humidity: 50%

Calibrated on: Channel 3 on 8G 5161 LE

PDA CALIBRATION FACTOR

10.0 g's/volt

R²: 0.999972 [No chip]

Ref Acc 1: LW503313! Cal on: 31Jan2024
10 g's/volt

Ref Acc 2: LW503311! Cal on: 31Jan2024
10 g's/volt


Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).

Operator: William Johnson


Signed

Calibration valid 2 years

From: 10Sep2024



Reference Velocity	S/N LW191303 Velocity
m/s	m/s
0.066	0.066
0.084	0.084
0.108	0.108
0.120	0.120
0.128	0.128
0.139	0.140
0.151	0.151
0.183	0.183
0.198	0.198
0.216	0.216
Maximum Acceleration: 23 g's	