
Certificate of Calibration

Pile Dynamics, Inc.
Transducer Model: PDI Strain Transducer
Serial Number: X339
PDI Gage Factor: 91.6
Mean Linear Correlation Coefficient: 0.999783

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-2132.5	-6.344	242.8	1.257
-1607.6	-4.698	590.6	2.332
-1095.8	-3.106	938.3	3.388
-715.2	-1.872	1273.0	4.431
-433.1	-0.886	1620.7	5.490
-111.5	0.153	2007.9	6.651

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: Paul Hartman

PDI Gage: X339
Calibration Date: 9/18/2024 12:34:46 PM