Certificate of Calibration
Pile Dynamics, Inc.

Calibrated by Pile Dynamics, Inc.

Serial No:	K4924	**Temperature:**	23.6 °C
Model:	KICHER PR	**Humidity:**	44%
Calibrated on:	Channel 3 on 8G 5161 LE		

Ref Acc 1:	78270! 971 g's/volt	**Cal on:**	11Jan2024
Ref Acc 2:	78268! 986 g's/volt	**Cal on:**	11Jan2024

Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).

PDA CALIBRATION FACTOR

338.7 mv/5000g
(67.7 µv/g)
R^2: 0.999985 [Chip programmed]

Operator: William Johnson

Signed

Calibration valid 2 years

From: 25Sep2024



Reference Velocity	S/N K4924 Velocity
m/s	m/s
0.930	0.927
1.464	1.466
1.841	1.843
2.216	2.214
2.586	2.585
2.936	2.931
3.206	3.205
3.557	3.556
3.933	3.929
4.315	4.325

Maximum Acceleration: 942 g's