Certificate of Calibration

Pile Dynamics, Inc.
Transducer Model: PDI Strain Transducer
Serial Number: Z682
PDI Gage Factor: 96.9
Mean Linear Correlation Coefficient: 0.999934

Table 1: Representative Calibration Data

Applied strain (με)	Gage Output (mV/Vexc)	Applied strain (με)	Gage Output (mV/Vexc)
-2027.6	-6.018	485.6	1.547
-1528.9	-4.517	912.1	2.804
-1047.7	-3.086	1273.0	3.865
-584.0	-1.695	1607.6	4.828
-229.7	-0.637	1994.8	5.928
85.3	0.316		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman Signature: Paul Hartman

PDI Gage: Z682
Calibration Date: OCT 2 5 2024