




Certificate of Calibration

Pile Dynamics, Inc.
 Transducer Model: PDI Strain Transducer
 Serial Number: V645
 PDI Gage Factor: 91.9
 Mean Linear Correlation Coefficient: 0.999588

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-2053.8	-6.511	255.9	0.903
-1522.3	-4.873	623.4	2.001
-1036.7	-3.351	964.6	3.067
-695.5	-2.218	1312.3	4.128
-406.8	-1.229	1686.4	5.215
-91.9	-0.208	1988.2	6.083

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature:

Paul Hartman

PDI Gage: V645

Calibration Date: 1/14/2025

