




Certificate of Calibration

Pile Dynamics, Inc.
 Transducer Model: PDI Strain Transducer
 Serial Number: X337
 PDI Gage Factor: 91.3
 Mean Linear Correlation Coefficient: 0.999801

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-2001.3	-6.411	406.8	1.375
-1627.3	-5.181	787.4	2.553
-1240.2	-3.921	1148.3	3.673
-833.3	-2.585	1522.3	4.796
-430.9	-1.288	1865.7	5.877
0.0	0.091	2001.3	6.257

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: _____

Paul Hartman

PDI Gage: X337

Calibration Date: 1/13/2025 2:14:48 PM

