



# Certificate of Calibration

TechSense

Transducer Model: PDI Strain Transducer

Serial Number: P498


PDI Gage Factor: 91.3

Mean Linear Correlation Coefficient: 0.999920

Table 1: Representative Calibration Data

| Applied strain ( $\mu\text{E}$ ) | Gage Output (mV/Vexc) | Applied strain ( $\mu\text{E}$ ) | Gage Output (mV/Vexc) |
|----------------------------------|-----------------------|----------------------------------|-----------------------|
| -1734.8                          | -5.818                | 395.9                            | 1.325                 |
| -1384.4                          | -4.608                | 539.8                            | 1.783                 |
| -1037.0                          | -3.445                | 859.0                            | 2.828                 |
| -673.0                           | -2.239                | 1225.7                           | 4.031                 |
| -315.5                           | -1.029                | 1583.1                           | 5.205                 |
| 40.2                             | 0.171                 | 1991.9                           | 6.509                 |

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

|                                     |                   |
|-------------------------------------|-------------------|
| PDI STCS Serial Number:             | 1003HA            |
| Firmware version number:            | 0.9.0.0           |
| Transducer Gage Length:             | 3 inches (76.2mm) |
| Excitation Voltage for Calibration: | 5.0 VDC           |


PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: RASENDU PATEL

Signature: \_\_\_\_\_

PDI Gage: P498

Calibration Date: 27-09-2024 17:02:37



## TechSense

49, Atmajyoti Nagar Society, Near Atmajyoti Ashram, Ellora Park, Vadodara - 390023

Website : [www.techsenseindia.com](http://www.techsenseindia.com) E-mail : [info@techsenseindia.com](mailto:info@techsenseindia.com)

Mobile.: +91- 8905510013 / +91 9327226069