



Certificate of Calibration

TechSense

Transducer Model: PDI Strain Transducer

Serial Number: U913


PDI Gage Factor: 137.4

Mean Linear Correlation Coefficient: 0.999910

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-1868.4	-4.158	526.6	1.116
-1410.2	-3.195	859.6	1.807
-1039.6	-2.386	1174.2	2.491
-728.6	-1.688	1480.4	3.152
-424.1	-0.986	1586.7	3.416
-101.2	-0.264	1706.6	3.687
208.5	0.430		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1003HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: RASENDU PATEL

Signature: _____

PDI Gage: U913

Calibration Date: 09-12-2024 16:39:27



TechSense

49, Atmajyoti Nagar Society, Near Atmajyoti Ashram, Ellora Park, Vadodara - 390023

Website : www.techsenseindia.com E-mail : info@techsenseindia.com

Mobile.: +91- 8905510013 / +91 9327226069