



Certificate of Calibration

Pile Dynamics, Inc.

Transducer Model: PDI Strain Transducer

Serial Number: W218


PDI Gage Factor: 90.8

Mean Linear Correlation Coefficient: 0.999649

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-1889.8	-6.358	308.4	1.043
-1496.1	-5.033	695.5	2.281
-1122.0	-3.762	1089.2	3.527
-728.3	-2.468	1450.1	4.655
-387.1	-1.316	1804.5	5.719
-52.5	-0.166	1975.1	6.256

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: _____

Paul Hartman

PDI Gage: W218

Calibration Date: 3/10/2025 7:27:10 AM

