


PILE DYNAMICS, INC.

6/20/86 CAL

GAGE NO. 102 D



INIT

6/20/86 CAL
GAGE NO. 1020

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	81.418	0.469
3	155.170	0.909
4	232.736	1.369
5	303.196	1.783
6	385.061	2.258
7	451.557	2.650
8	528.175	3.098
9	599.546	3.519
10	674.675	3.961
11	752.521	4.428

GAGE CALIBRATION FACTOR= 170.170 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = 0.6262 CORRELATION COEFFICIENT = 1.0000

VALUES ARE AVERAGED