


Pile Dynamics, Inc.

Transducer 104M



PDA Cal Factor = 113.5 ME/V
 Shunt (60.4KΩ) = 5.0 V
 General Factor = 394.0 ME/mV/V
 Traceable to N.I.S.T.

Strain ME	Gage V
0	.00
57	.53
169	1.51
223	2.01
314	2.80
387	3.45
474	4.20
554	4.91
623	5.51
721	6.37
774	6.84

Calibrated by: Jim Stachoff

Calibrated on: 09-May-95