


Pile Dynamics, Inc. Transducer 105M



PDA Cal Factor = 109.2 ME/V
 Shunt (60.4KΩ) = 5.0 V
 General Factor = 379.3 ME/mV/V
 Traceable to N.I.S.T.

Strain ME	Gage V
0	.00
62	.58
159	1.48
225	2.09
291	2.69
364	3.37
452	4.17
525	4.83
612	5.61
680	6.23
762	6.99

Calibrated by: Jim Stetefeld

Calibrated on: 09-May-95