


PILE DYNAMICS, INC.

GAGE NO. 111 W

07-Sep-89



PDA Gage Factor = 116.5 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS

General Gage Factor = 404.4 ME/mV/V
 Statistical Offset = -3.3

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	89.6	.80
3	164.9	1.46
4	232.1	2.03
5	308.5	2.68
6	389.7	3.38
7	469.4	4.07
8	539.6	4.67
9	613.3	5.29
10	691.4	5.96
11	766.6	6.60