


PILE DYNAMICS, INC.

GAGE NO. 113 W

28-Sep-89



PDA Gage Factor = 105.1 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS

General Gage Factor = 365.0 ME/mV/V
 Statistical Offset = -2.2

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	78.2	.75
3	156.4	1.51
4	229.9	2.21
5	316.6	3.06
6	390.9	3.76
7	465.6	4.46
8	535.9	5.12
9	617.0	5.88
10	689.5	6.58
11	763.4	7.27