


PILE DYNAMICS, INC.

GAGE NO. 114 W

28-Sep-89



PDA Gage Factor = 107.3 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS

General Gage Factor = 372.5 ME/mV/V
 Statistical Offset = -4.9

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	78.2	.74
3	156.4	1.50
4	229.9	2.20
5	316.6	3.03
6	390.9	3.72
7	465.6	4.41
8	535.9	5.05
9	617.0	5.78
10	689.5	6.45
11	763.4	7.11