


PILE DYNAMICS, INC.

GAGE NO. 115,D

27-JUN-87



Gage Calibration Factor = 112.9 ME/V
 Gage Ref. Shunt (50.4K) = 5.0 VOLTS
 Offset = -4.1

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	80.4	.75
3	157.2	1.42
4	234.3	2.11
5	309.8	2.81
6	388.3	3.50
7	462.8	4.15
8	539.5	4.84
9	618.5	5.50
10	692.8	6.15
11	769.5	6.81