


PILE DYNAMICS, INC.

GAGE NO. 116,D

27-Jun-87



Gage Calibration Factor = 108.9 ME/V
 Gage Ref. Shunt (50.4K) = 5.0 VOLTS
 Offset = -4.5

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	80.4	.77
3	157.2	1.47
4	234.3	2.20
5	309.8	2.91
6	388.3	3.63
7	462.8	4.32
8	539.5	5.01
9	618.5	5.69
10	692.8	6.39
11	769.6	7.07