


PILE DYNAMICS, INC.

GAGE NO. 116 W

28-Sep-89



PDA Gage Factor = 118.2 ME/V
 Gage Ref. Shunt (60.4K0) = 5.0 VOLTS

General Gage Factor = 410.3 ME/mV/V
 Statistical Offset = -5.9

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	85.7	.75
3	152.6	1.34
4	230.5	2.01
5	312.5	2.73
6	384.1	3.34
7	469.2	4.05
8	537.6	4.62
9	613.3	5.25
10	691.8	5.90
11	765.1	6.49