


PILE DYNAMICS, INC.

GAGE NO. 117D

25-May-88



Gage Calibration Factor = 111.0 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -3.1

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	76.6	.70
3	160.6	1.48
4	246.0	2.26
5	322.2	2.95
6	395.4	3.61
7	479.9	4.39
8	537.9	4.90
9	617.9	5.61
10	698.2	6.30
11	765.4	6.88