


PILE DYNAMICS, INC.

GAGE NO. 117 W

29-Sep-89



PDA Gage Factor = 108.6 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS
 General Gage Factor = 376.9 ME/mV/V
 Statistical Offset = -4.6

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	83.1	.79
3	152.1	1.45
4	231.4	2.19
5	309.4	2.92
6	381.4	3.59
7	467.9	4.36
8	535.1	4.98
9	610.0	5.67
10	690.6	6.39
11	766.0	7.06