


PILE DYNAMICS, INC.

GAGE NO. 118 W

29-Sep-89



PDA Gage Factor = 107.7 ME/V
 Gage Ref. Shunt (60.4K0) = 5.0 VOLTS
 General Gage Factor = 373.9 ME/mV/V
 Statistical Offset = -3.7

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	83.1	.78
3	152.1	1.45
4	231.4	2.22
5	309.4	2.94
6	381.4	3.61
7	467.9	4.40
8	535.1	5.01
9	610.0	5.70
10	690.6	6.43
11	766.0	7.10