


PILE DYNAMICS, INC.

GAGE NO. 121 D

19-Feb-92



PDA Gage Factor = 104.1 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS

General Gage Factor = 361.4 ME/mV/V
 Statistical Offset = -3.6

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	78.1	.77
3	158.9	1.56
4	237.0	2.32
5	308.6	3.03
6	368.9	3.79
7	463.6	4.52
8	538.5	5.23
9	613.6	5.93
10	692.9	6.67
11	765.4	7.35