


PILE DYNAMICS, INC.

GAGE 122W

05-Mar-92



PDA Cal Factor = 109.7 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 380.8 ME/mV/V

Strain ME	Gage Volts
0.	.00
75.	.70
154.	1.41
230.	2.12
306.	2.81
384.	3.53
458.	4.19
538.	4.91
612.	5.59
689.	6.29
769.	7.02