


PILE DYNAMICS, INC.

GAGE NO. 124D

28-JUN-87



Gage Calibration Factor = 110.6 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS
 Offset = -6.8

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	79.4	.76
3	158.5	1.51
4	232.9	2.19
5	310.5	2.91
6	386.1	3.58
7	463.0	4.28
8	539.1	4.95
9	616.1	5.64
10	690.8	6.28
11	767.9	6.96