


PILE DYNAMICS, INC.

GAGE NO. 1250

01-Sep-88



Gage Calibration Factor = 110.1 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -3.7

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	79.5	.74
3	157.4	1.46
4	237.7	2.21
5	312.9	2.91
6	397.6	3.68
7	466.4	4.28
8	545.0	4.98
9	613.8	5.60
10	692.3	6.31
11	772.9	7.02