


PILE DYNAMICS, INC.

GAGE NO. 126D

19-JUL-87



Gage Calibration Factor = 116.4 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -3.0

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	74.7	.64
3	150.9	1.33
4	229.4	2.01
5	304.6	2.67
6	382.2	3.34
7	459.0	3.99
8	535.3	4.64
9	613.0	5.29
10	691.0	5.95
11	766.6	6.59