


PILE DYNAMICS, INC.

GAGE NO. 127D

19-JUL-87



Gage Calibration Factor = 116.7 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -5.2

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	75.8	.72
3	154.7	1.41
4	231.7	2.03
5	308.6	2.67
6	384.7	3.35
7	460.2	4.01
8	540.8	4.62
9	615.0	5.36
10	690.9	6.00
11	768.9	6.59