


PILE DYNAMICS, INC.

GAGE 153W

10-Aug-92



PDA Cal Factor = 191.5 ME/V
 Ref. Shunt (60.4K) = 5.0 V

General Gage Factor= 665.0 ME/mV/V

Strain ME	Gage Volts
0.	.00
97.	.50
180.	.93
255.	1.33
344.	1.80
432.	2.25
518.	2.69
616.	3.21
692.	3.61
783.	4.09
867.	4.54