


PILE DYNAMICS, INC.

GAGE NO. 154 D

05-Oct-89



PDA Gage Factor = 102.5 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS

General Gage Factor = 355.8 ME/mV/V
 Statistical Offset = -1.4

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	76.5	.74
3	150.0	1.48
4	232.1	2.29
5	306.4	3.02
6	385.6	3.80
7	456.7	4.48
8	540.1	5.29
9	614.3	6.02
10	685.5	6.69
11	759.5	7.41