


PILE DYNAMICS, INC.

GAGE NO. 1550

08-Aug-87



Gage Calibration Factor = 118.6 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -4.5

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	78.1	.68
3	154.5	1.33
4	231.1	2.00
5	309.3	2.68
6	386.1	3.31
7	462.5	3.96
8	538.1	4.59
9	616.6	5.24
10	692.2	5.84
11	769.9	6.50