


PILE DYNAMICS, INC.

GAGE NO. 157D

23-Aug-87



Gage Calibration Factor = 114.4 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -6.3

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	79.5	.74
3	156.0	1.43
4	234.2	2.12
5	309.3	2.80
6	384.6	3.45
7	461.3	4.11
8	537.5	4.77
9	615.2	5.43
10	691.6	6.07
11	770.8	6.76