




Pile Dynamics, Inc.

Pile Dynamics, Inc.

Transducer

158W

Strain ($\mu\epsilon$)



Traceable to N.I.S.T.

Strain Transducer Calibrator System 2004 Version 1.2

PDA Cal Factor

194.1 $\mu\epsilon/V$

Applied Strain ($\mu\epsilon$)	Transducer Output (Volts)
0	0.00
99	0.53
201	1.06
301	1.59
391	2.05
500	2.61
575	2.99
677	3.51
785	4.06
882	4.56
994	5.13

Shunt (60.4 K Ω) 5.0 V
General Factor 674.1 $\mu\epsilon/mV/V$

Calibrated by: John R. Rasmussen
Calibrated on: 12-Nov-2004